

Pratik Patel

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Monday, November 25, 2024 12:49 PM
To: Sohil Jodhani; Nimisha Pandya; Mohammad Ahmed
Cc: R5 - R5RSCC@epa.gov (R5RSCC@epa.gov)
Subject: REGION 5 | CASE 51834 | LAB ACE | SDG E27M3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: E27M7.pdf; Internal Form.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon ACE,

Regarding the laboratory's reported issue for Case 51834/SDG E27M3, please see the Region's input below and let me know if you have any questions.

LAB: "Lab has received water samples for VOA analysis. Lab has analyzed undiluted VOA analysis for the sample E27M7. Sample found positive with extremely elevated target analytes from calibration range and required dilution as you can see attached form-1 for your reference. Due to matrix interference, sample has internal standard recovery outside the QC limits therefore lab would like to confirm that lab will report undiluted VOA analysis with internal standard failure and further dilution in final electronic deliverables.

Please see attachment for your reference."

REGION: "The lab's proposal is acceptable."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in

error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>

Sent: Monday, November 25, 2024 12:15 PM

To: Irving, Madison <Madison.Irving@gdit.com>

Subject: RE: REGION 5 | CASE 51834 | LAB ACE | SDG E27M3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Madison,

The lab's proposal is acceptable.

Thanks,
Amanda

Amanda Wroble, Ph.D.

Metals Chemist & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator

USEPA Region 5 LSASD/AIMB

Chicago Regional Laboratory

536 S. Clark St. LAB-10C, 10th Floor

Chicago, IL 60605

(312) 353-0375

From: Irving, Madison <Madison.Irving@gdit.com>

Sent: Monday, November 25, 2024 10:34 AM

To: R5RSCC <R5RSCC@epa.gov>

Subject: REGION 5 | CASE 51834 | LAB ACE | SDG E27M3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 5,

Please advise on the laboratory's reported issue below for Case 51834/SDG E27M3.

LAB: "Lab has received water samples for VOA analysis. Lab has analyzed undiluted VOA analysis for the sample E27M7. Sample found positive with extremely elevated target analytes from calibration range and required dilution as you can see attached form-1 for your reference. Due to matrix interference, sample has internal standard recovery outside the QC limits therefore lab would like to confirm that lab will report undiluted VOA analysis with internal standard failure and further dilution in final electronic deliverables.

Please see attachment for your reference.”

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport
Sent: Monday, November 25, 2024 11:12 AM
To: 'Sohil Jodhani' <Sohil.Jodhani@alliancetg.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 5 | CASE 51834 | LAB ACE | SDG E27M3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASsupport@gdit.com

GDIT Federal Civilian Division
Under contract to the EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure,

copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Sent: Monday, November 25, 2024 10:08 AM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 5 | CASE 51834 | LAB ACE | SDG E27M3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is seeing this email with regards to case 51834 and SDG E27M3.

Lab has received water samples for VOA analysis. Lab has analyzed undiluted VOA analysis for the sample E27M7. Sample found positive with extremely elevated target analytes from calibration range and required dilution as you can see attached form-1 for your reference. Due to matrix interference, sample has internal standard recovery outside the QC limits therefore lab would like to confirm that lab will report undiluted VOA analysis with internal standard failure and further dilution in final electronic deliverables.

Please see attachment for your reference.

Thanks & Regards,



Sohil Jodhani

QA/QC Director

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3152

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com

