

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51874 MA No. : SDG No.: GCP96
 Analytical Method: ICP-MS ICSA Source: EPA-0803
 Instrument ID: P7 ICSB Source: EPA-0803
 Run Batch: LB133416
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	100000	100000	93000	93	90000	90
Antimony	1.5	22	1.1	76	21	97
Arsenic	0.1	19	0.32	320	20	106
Barium	1.2	22	1.5	122	21	97
Beryllium	0	19	0.3		20	107
Cadmium	0.7	20	0.39	56	21	102
Calcium	100000	100000	100000	102	100000	100
Chromium	21	40	20	95	39	98
Cobalt	1	20	1.3	125	21	104
Copper	8	25	7.1	89	26	102
Iron	100000	100000	100000	104	100000	103
Lead	4	25	4.6	115	25	98
Magnesium	100000	100000	100000	101	98000	98
Manganese	7	27	8.0	114	27	101
Nickel	6	24	5.3	88	25	103
Potassium	100000	100000	100000	101	99000	99
Selenium	0.3	19	0.4	133	19	101
Silver	0	18	0.03		19	107
Sodium	100000	100000	97000	97	96000	96
Thallium	0	21	0.12		21	98
Vanadium	0.5	19	0.17	34	19	101
Zinc	11	29	11	99	30	105