

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133398

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/11/2024	16:25
2	CAL2	1	Water	NA	NA	20.3	7.00	11/11/2024	16:26
3	CAL3	1	Water	NA	NA	20.2	10.02	11/11/2024	16:27
4	ICV	1	Water	NA	NA	20.3	6.98	11/11/2024	16:30
5	CCV1	1	Water	NA	NA	20.3	2.02	11/11/2024	16:35
6	P4799-02	1	Solid	20.02	20	24.2	11.67	11/11/2024	16:40
7	P4799-02DUP	1	Solid	20.03	20	24.3	11.68	11/11/2024	16:41
8	P4799-06	1	Solid	20.02	20	24.3	11.91	11/11/2024	16:42
9	P4799-10	1	Solid	20.03	20	24.4	12.04	11/11/2024	16:45
10	P4799-14	1	Solid	20.02	20	24.3	12.03	11/11/2024	16:50
11	P4799-18	1	Solid	20.03	20	24.4	12.27	11/11/2024	16:55
12	CCV2	1	Water	NA	NA	20.3	12.01	11/11/2024	17:00

WORKLIST(Hardcopy Internal Chain)

VB133398

WorkList Name : ph s p4799

WorkList ID : 185340

Department : Wet-Chemistry

Date : 11-11-2024 16:29:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4799-02	WC-TA2-02-C	Solid	pH	Cool 4 deg C	ENTA05	L31	10/30/2024	9045D
P4799-06	WC-TA2-03-C	Solid	pH	Cool 4 deg C	ENTA05	L31	11/01/2024	9045D
P4799-10	WC-TA1-01-C	Solid	pH	Cool 4 deg C	ENTA05	L31	11/04/2024	9045D
P4799-14	WC-TA1-02-C	Solid	pH	Cool 4 deg C	ENTA05	L31	11/05/2024	9045D
P4799-18	WC-TA1-03-C	Solid	pH	Cool 4 deg C	ENTA05	L31	11/06/2024	9045D

Date/Time 11-11-24 16:10
Raw Sample Received by: JO WLE
Raw Sample Relinquished by: RS (EST-1060)

Date/Time 11-11-24
Raw Sample Received by: RS (EST-1060)
Raw Sample Relinquished by: JO WLE