



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133397

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	6.98	7	100	90-110	11/11/2024
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	11/11/2024
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	11/11/2024
Sample ID: Corrosivity	CCV3	pH	2.02	2.00	101	90-110	11/11/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133398

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	6.98	7	100	90-110	11/11/2024
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	11/11/2024
Sample ID:	CCV2						
pH		pH	12.01	12.00	100	90-110	11/11/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133401

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.097	0.099	98	85-115	11/11/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/11/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/11/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/11/2024
Sample ID: CCV4 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	11/11/2024
Sample ID: CCV5 Reactive Cyanide	mg/L	0.27	0.25	108	90-110	11/11/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133425

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	0.98	1	98	90-110	11/13/2024
Sample ID: CCV1 ASTM Ammonia	mg/L	1	1	100	90-110	11/13/2024
Sample ID: CCV2 ASTM Ammonia	mg/L	1	1	100	90-110	11/13/2024
Sample ID: CCV3 ASTM Ammonia	mg/L	1	1	100	90-110	11/13/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133460

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
ASTM COD		mg/L	51.341	50	103	95-105	10/14/2024
Sample ID:	CCV1						
ASTM COD		mg/L	50.336	50	101	95-105	11/14/2024
Sample ID:	CCV2						
ASTM COD		mg/L	50.336	50	101	95-105	11/14/2024