

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133404

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133404BS							
ASTM Oil and Grease	mg/L	20.0	16.7		84	1	78-114	11/12/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133422

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133422BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	11/13/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133460

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133460BS							
ASTM COD	mg/L	50	49.3		99	1	90-110	11/14/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4799

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133425

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164917BS							
ASTM Ammonia	mg/L	1	1.00		100	1	90-110	11/13/2024