


LB133446

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	39	0.000	39	0				100
0.2	0.200	0.247	0.047	866	0.0 %	866					24
2.5	2.500	2.409	-0.091	8865	0.0 %	8865					4
5.0	5.000	4.993	-0.007	18424	0.0 %	18424					0
7.5	7.500	7.516	0.016	27757	0.0 %	27757					0
10.0	10.000	10.012	0.012	36991	0.0 %	36991					0

LB133446 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Nov 2024 11:37:45

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7470A	PPB	14 Nov 2024 11:59:16	S	Std
0.2 - 1	50.2	866	-	0.2000	7470A	PPB	14 Nov 2024 12:01:33	S	Std
2.5 - 1	52.5	8865	-	2.5000	7470A	PPB	14 Nov 2024 12:03:49	S	Std
5.0 - 1	55	18424	-	5.0000	7470A	PPB	14 Nov 2024 12:08:55	S	Std
7.5 - 1	57.5	27757	-	7.5000	7470A	PPB	14 Nov 2024 12:17:14	S	Std
10.0 - 1	57.5	36991	-	10.0000	7470A	PPB	14 Nov 2024 12:25:14	S	Std
ICV85 - 1	ICV85	14990	4.0647	-	7470A	PPB	14 Nov 2024 12:38:23	U	SMPL
ICB85 - 1	ICB85	-311	-0.0715	-	7470A	PPB	14 Nov 2024 12:40:37	U	SMPL
CCV83 - 1	CCV83	18361	4.9760	-	7470A	PPB	14 Nov 2024 12:45:57	U	SMPL
CCB83 - 1	CCB83	-337	-0.0785	-	7470A	PPB	14 Nov 2024 12:48:11	U	SMPL
CRA - 1	CRA	643	0.1864	-	7470A	PPB	14 Nov 2024 12:50:28	U	SMPL
HighStd - 1	HighStd	36451	9.8662	-	7470A	PPB	14 Nov 2024 12:52:43	U	SMPL
ChkStd - 1	ChkStd	23225	6.2909	-	7470A	PPB	14 Nov 2024 12:55:00	U	SMPL
PB164955BS - 1	LCSW	15348	4.1615	-	7470A	PPB	14 Nov 2024 12:59:36	U	SMPL
P4799-03 - 1	WC-TA2-02-C	1648	0.4581	-	7470A	PPB	14 Nov 2024 13:01:53	U	SMPL
P4799-03DUP - 1	WC-TA2-02-CDUP	1984	0.5489	-	7470A	PPB	14 Nov 2024 13:04:11	U	SMPL
P4799-03MS - 1	WC-TA2-02-CMS	9811	2.6647	-	7470A	PPB	14 Nov 2024 13:06:28	U	SMPL
P4799-03MSD - 1	WC-TA2-02-CMSD	10442	2.8353	-	7470A	PPB	14 Nov 2024 13:08:45	U	SMPL
P4799-07 - 1	WC-TA2-03-C	341	0.1048	-	7470A	PPB	14 Nov 2024 13:11:02	U	SMPL
CCV84 - 1	CCV84	18867	5.1128	-	7470A	PPB	14 Nov 2024 13:17:17	U	SMPL
CCB84 - 1	CCB84	-405	-0.0969	-	7470A	PPB	14 Nov 2024 13:19:32	U	SMPL
P4799-11 - 1	WC-TA1-01-C	990	0.2802	-	7470A	PPB	14 Nov 2024 13:23:11	U	SMPL
P4799-15 - 1	WC-TA1-02-C	725	0.2086	-	7470A	PPB	14 Nov 2024 13:25:25	U	SMPL
P4799-19 - 1	WC-TA1-03-C	703	0.2026	-	7470A	PPB	14 Nov 2024 13:27:40	U	SMPL
P4807-04 - 1	TP-7	-25	0.0058	-	7470A	PPB	14 Nov 2024 13:29:56	U	SMPL
P4807-08 - 1	BF-F14	-57	-0.0028	-	7470A	PPB	14 Nov 2024 13:32:11	U	SMPL
P4807-12 - 1	BF-F13	-75	-0.0077	-	7470A	PPB	14 Nov 2024 13:34:27	U	SMPL
P4807-16 - 1	TP-6	-69	-0.0061	-	7470A	PPB	14 Nov 2024 13:36:44	U	SMPL
P4809-04 - 1	MH-5	1460	0.4073	-	7470A	PPB	14 Nov 2024 13:39:00	U	SMPL
P4813-02 - 1	417	2570	0.7073	-	7470A	PPB	14 Nov 2024 13:41:17	U	SMPL
P4813-03 - 1	418	2613	0.7189	-	7470A	PPB	14 Nov 2024 13:43:33	U	SMPL
CCV85 - 1	CCV85	17946	4.8638	-	7470A	PPB	14 Nov 2024 13:51:00	U	SMPL
CCB85 - 1	CCB85	-288	-0.0653	-	7470A	PPB	14 Nov 2024 13:53:15	U	SMPL
P4813-05 - 1	420	3319	0.9098	-	7470A	PPB	14 Nov 2024 13:55:32	U	SMPL
PB164956BL - 1	PBW	-172	-0.0339	-	7470A	PPB	14 Nov 2024 13:57:47	U	SMPL
PB164956BS - 1	LCSW	15331	4.1569	-	7470A	PPB	14 Nov 2024 14:00:02	U	SMPL
P4812-03 - 1	270	-273	-0.0612	-	7470A	PPB	14 Nov 2024 14:02:18	U	SMPL
P4813-01 - 1	416	88	0.0364	-	7470A	PPB	14 Nov 2024 14:04:35	U	SMPL
P4813-04 - 1	419	-107	-0.0163	-	7470A	PPB	14 Nov 2024 14:06:51	U	SMPL
P4813-04DUP - 1	419DUP	-42	0.0012	-	7470A	PPB	14 Nov 2024 14:09:08	U	SMPL
P4813-04MS - 1	419MS	15398	4.1750	-	7470A	PPB	14 Nov 2024 14:11:24	U	SMPL
P4813-04MSD - 1	419MSD	14143	3.8358	-	7470A	PPB	14 Nov 2024 14:13:41	U	SMPL
PB164975BL - 1	PBW	-339	-0.0791	-	7470A	PPB	14 Nov 2024 14:15:58	U	SMPL
CCV86 - 1	CCV86	20023	5.4253	-	7470A	PPB	14 Nov 2024 14:26:08	U	SMPL
CCB86 - 1	CCB86	-356	-0.0837	-	7470A	PPB	14 Nov 2024 14:28:23	U	SMPL
PB164975BS - 1	LCSW	15574	4.2226	-	7470A	PPB	14 Nov 2024 14:30:40	U	SMPL
P4834-01 - 1	SLP-1-WATER	-167	-0.0326	-	7470A	PPB	14 Nov 2024 14:32:56	U	SMPL
P4834-01DUP - 1	SLP-1-WATERDUP	-44	0.0007	-	7470A	PPB	14 Nov 2024 14:35:13	U	SMPL
P4834-01MS - 1	SLP-1-WATERMS	14580	3.9539	-	7470A	PPB	14 Nov 2024 14:37:29	U	SMPL
P4834-01MSD - 1	SLP-1-WATERMSD	13603	3.6898	-	7470A	PPB	14 Nov 2024 14:39:44	U	SMPL
P4835-01 - 1	345-2-WATER	10916	2.9634	-	7470A	PPB	14 Nov 2024 14:42:01	U	SMPL
P4836-01 - 1	T3-WATER	-169	-0.0331	-	7470A	PPB	14 Nov 2024 14:44:19	U	SMPL
P4837-01 - 1	3TRK-WATER	-104	-0.0155	-	7470A	PPB	14 Nov 2024 14:46:39	U	SMPL
P4838-01 - 1	T1-WATER	14194	3.8496	-	7470A	PPB	14 Nov 2024 14:48:56	U	SMPL
P4843-01 - 1	SW-WTS-01	1289	0.3610	-	7470A	PPB	14 Nov 2024 14:51:11	U	SMPL
CCV87 - 1	CCV87	20049	5.4323	-	7470A	PPB	14 Nov 2024 14:53:29	U	SMPL
CCB87 - 1	CCB87	-117	-0.0190	-	7470A	PPB	14 Nov 2024 14:55:45	U	SMPL
PB164880TB - 1	PB164880TB	88	0.0364	-	7470A	PPB	14 Nov 2024 15:00:10	U	SMPL
PB164890TB - 1	PB164890TB	92	0.0375	-	7470A	PPB	14 Nov 2024 15:02:26	U	SMPL
P4799-03LX5 - 1		493	0.1459	-	7470A	PPB	14 Nov 2024 15:04:41	U	SMPL
P4799-03A - 1		15379	4.1699	-	7470A	PPB	14 Nov 2024 15:06:56	U	SMPL
P4813-04LX5 - 1		-133	-0.0234	-	7470A	PPB	14 Nov 2024 15:09:12	U	SMPL
P4813-04A - 1		16928	4.5886	-	7470A	PPB	14 Nov 2024 15:11:29	U	SMPL
P4834-01LX5 - 1		-124	-0.0209	-	7470A	PPB	14 Nov 2024 15:13:45	U	SMPL
P4834-01A - 1		21695	5.8773	-	7470A	PPB	14 Nov 2024 15:16:03	U	SMPL
PB164955BL - 1	PBW	-157	-0.0299	-	7470A	PPB	14 Nov 2024 15:18:21	U	SMPL
CCV88 - 1	CCV88	16806	4.5557	-	7470A	PPB	14 Nov 2024 15:38:19	U	SMPL
CCB88 - 1	CCB88	-207	-0.0434	-	7470A	PPB	14 Nov 2024 15:40:36	U	SMPL