



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** P4799
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4799 **SAS No.:** P4799
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	93	40 - 160	CV	11/14/2024	12:50	LB133446
CRI01	Arsenic	20.6	20.0	103	40 - 160	P	11/14/2024	14:40	LB133457
	Barium	95.9	100	96	40 - 160	P	11/14/2024	14:40	LB133457
	Cadmium	6.51	6.0	108	40 - 160	P	11/14/2024	14:40	LB133457
	Chromium	9.55	10.0	96	40 - 160	P	11/14/2024	14:40	LB133457
	Copper	21.3	20.0	107	40 - 160	P	11/14/2024	14:40	LB133457
	Lead	11.8	12.0	98	40 - 160	P	11/14/2024	14:40	LB133457
	Nickel	36.7	40.0	92	40 - 160	P	11/14/2024	14:40	LB133457
	Selenium	17.4	20.0	87	40 - 160	P	11/14/2024	14:40	LB133457
	Silver	9.72	10.0	97	40 - 160	P	11/14/2024	14:40	LB133457
	Zinc	41.6	40.0	104	40 - 160	P	11/14/2024	14:40	LB133457