

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** P4799
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4799 **SAS No.:** P4799
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-1.10			-20	20	11/14/2024	14:54	LB133457
	Barium	3.24	6.0	54	-94	106	11/14/2024	14:54	LB133457
	Cadmium	6.64	1.0	664	-5	7	11/14/2024	14:54	LB133457
	Chromium	55.5	52.0	107	42	62	11/14/2024	14:54	LB133457
	Copper	-4.72	2.0	236	-18	22	11/14/2024	14:54	LB133457
	Lead	8.61			-12	12	11/14/2024	14:54	LB133457
	Nickel	2.90	2.0	145	-38	42	11/14/2024	14:54	LB133457
	Selenium	-16.4			-20	20	11/14/2024	14:54	LB133457
	Silver	2.63			-10	10	11/14/2024	14:54	LB133457
	Zinc	6.48			-40	40	11/14/2024	14:54	LB133457
ICSAB01	Arsenic	117	104	112	88.4	120	11/14/2024	15:18	LB133457
	Barium	535	537	100	437	637	11/14/2024	15:18	LB133457
	Cadmium	1020	972	105	826	1120	11/14/2024	15:18	LB133457
	Chromium	568	542	105	460	624	11/14/2024	15:18	LB133457
	Copper	493	511	96	434	588	11/14/2024	15:18	LB133457
	Lead	56.9	49.0	116	37	61	11/14/2024	15:18	LB133457
	Nickel	1020	954	107	810	1100	11/14/2024	15:18	LB133457
	Selenium	35.8	46.0	78	26	66	11/14/2024	15:18	LB133457
	Silver	206	201	102	170	232	11/14/2024	15:18	LB133457
	Zinc	852	952	90	809	1095	11/14/2024	15:18	LB133457