

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 09:27 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1016-1       | (1) | 5.53       | 5.53      | 5.43                   | 5.63  | 0.00       |
| Aroclor-1016-2       | (2) | 5.55       | 5.55      | 5.45                   | 5.65  | 0.00       |
| Aroclor-1016-3       | (3) | 5.61       | 5.61      | 5.51                   | 5.71  | 0.00       |
| Aroclor-1016-4       | (4) | 5.71       | 5.71      | 5.61                   | 5.81  | 0.00       |
| Aroclor-1016-5       | (5) | 6.01       | 6.01      | 5.91                   | 6.11  | 0.00       |
| Aroclor-1260-1       | (1) | 7.14       | 7.14      | 7.04                   | 7.24  | 0.00       |
| Aroclor-1260-2       | (2) | 7.40       | 7.40      | 7.30                   | 7.50  | 0.00       |
| Aroclor-1260-3       | (3) | 7.76       | 7.76      | 7.66                   | 7.86  | 0.00       |
| Aroclor-1260-4       | (4) | 7.98       | 7.98      | 7.88                   | 8.08  | 0.00       |
| Aroclor-1260-5       | (5) | 8.30       | 8.30      | 8.20                   | 8.40  | 0.00       |
| Tetrachloro-m-xylene |     | 4.37       | 4.37      | 4.27                   | 4.47  | 0.00       |
| Decachlorobiphenyl   |     | 10.09      | 10.09     | 9.99                   | 10.19 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 09:27 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.72       | 4.72      | 4.62      | 4.82 | 0.00       |
| Aroclor-1016-2       | (2) | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1016-3       | (3) | 4.92       | 4.92      | 4.82      | 5.02 | 0.00       |
| Aroclor-1016-4       | (4) | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1016-5       | (5) | 5.17       | 5.17      | 5.07      | 5.27 | 0.00       |
| Aroclor-1260-1       | (1) | 6.20       | 6.20      | 6.10      | 6.30 | 0.00       |
| Aroclor-1260-2       | (2) | 6.39       | 6.39      | 6.29      | 6.49 | 0.00       |
| Aroclor-1260-3       | (3) | 6.55       | 6.55      | 6.45      | 6.65 | 0.00       |
| Aroclor-1260-4       | (4) | 7.02       | 7.02      | 6.92      | 7.12 | 0.00       |
| Aroclor-1260-5       | (5) | 7.26       | 7.26      | 7.16      | 7.36 | 0.00       |
| Tetrachloro-m-xylene |     | 3.64       | 3.64      | 3.54      | 3.74 | 0.00       |
| Decachlorobiphenyl   |     | 8.64       | 8.64      | 8.54      | 8.74 | 0.00       |