

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 17:09 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.53	5.53	5.43	5.63	0.00
Aroclor-1242-2 (2)	5.55	5.55	5.45	5.65	0.00
Aroclor-1242-3 (3)	5.61	5.61	5.51	5.71	0.00
Aroclor-1242-4 (4)	5.71	5.71	5.61	5.81	0.00
Aroclor-1242-5 (5)	6.45	6.45	6.35	6.55	0.00
Tetrachloro-m-xylene	4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl	10.09	10.09	9.99	10.19	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 17:09 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1242-2 (2)	4.74	4.74	4.64	4.84	0.00
Aroclor-1242-3 (3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1242-4 (4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1242-5 (5)	5.52	5.52	5.42	5.62	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.64	8.64	8.54	8.74	0.00