

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/13/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 14:26 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.52	5.53	5.43	5.63	0.01
Aroclor-1248-2 (2)	5.80	5.80	5.70	5.90	0.00
Aroclor-1248-3 (3)	6.00	6.00	5.90	6.10	0.00
Aroclor-1248-4 (4)	6.40	6.41	6.31	6.51	0.01
Aroclor-1248-5 (5)	6.44	6.45	6.35	6.55	0.01
Tetrachloro-m-xylene	4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl	10.08	10.09	9.99	10.19	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/13/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 14:26 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1248-2 (2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1248-3 (3)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-4 (4)	5.17	5.17	5.07	5.27	0.00
Aroclor-1248-5 (5)	5.56	5.56	5.46	5.66	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01