

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/13/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 19:21 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.38	6.38	6.28	6.48	0.00
Aroclor-1254-2	(2)	6.60	6.60	6.50	6.70	0.00
Aroclor-1254-3	(3)	6.97	6.97	6.87	7.07	0.00
Aroclor-1254-4	(4)	7.25	7.26	7.16	7.36	0.01
Aroclor-1254-5	(5)	7.67	7.68	7.58	7.78	0.01
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.08	10.09	9.99	10.19	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

Continuing Calib Date: 11/13/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 19:21 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1254-2	(2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1254-3	(3)	6.07	6.07	5.97	6.17	0.00
Aroclor-1254-4	(4)	6.30	6.30	6.20	6.40	0.00
Aroclor-1254-5	(5)	6.72	6.72	6.62	6.82	0.00
Tetrachloro-m-xylene		3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01