284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** P4799 **SAS No.:** P4799 **SDG NO.:** P4799

Continuing Calib Date: 11/13/2024 **Initial Calibration Date(s):** 11/11/2024 11/12/2024

Continuing Calib Time: 23:37 **Initial Calibration Time(s):** 17:30 00:52

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1248-1	(1)	5.53	5.53	5.43	5.63	0.00
Aroclor-1248-2	(2)	5.80	5.80	5.70	5.90	0.00
Aroclor-1248-3	(3)	6.01	6.00	5.90	6.10	-0.01
Aroclor-1248-4	(4)	6.41	6.41	6.31	6.51	0.00
Aroclor-1248-5	(5)	6.44	6.45	6.35	6.55	0.01
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.08	10.09	9.99	10.19	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** P4799 **SAS No.:** P4799 **SDG NO.:** P4799

Continuing Calib Date: 11/13/2024 **Initial Calibration Date(s):** 11/11/2024 11/12/2024

Continuing Calib Time: 23:37 **Initial Calibration Time(s):** 17:30 00:52

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1248-1	(1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1248-2	(2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1248-3	(3)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-4	(4)	5.17	5.17	5.07	5.27	0.00
Aroclor-1248-5	(5)	5.56	5.56	5.46	5.66	0.00
Tetrachloro-m-xylene		3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01