284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

## CALIBRATION VERIFICATION SUMMARY

**Contract:** ENTA05

**Lab Code:** CHEM **Case No.:** P4799 **SAS No.:** P4799 **SDG NO.:** P4799

**Continuing Calib Date:** 11/13/2024 **Initial Calibration Date(s):** 11/11/2024 11/12/2024

**Continuing Calib Time:** 23:54 **Initial Calibration Time(s):** 17:30 00:52

**GC Column:** ZB-MR1 **ID:** 0.32 **(mm)**

| COMPOUND | | CCAL RT | AVG RT | RT WINDOW FROM | RT WINDOW TO | DIFF RT |
|---|---|---|---|---|---|---|
| Aroclor-1254-1 | (1) | 6.38 | 6.38 | 6.28 | 6.48 | 0.00 |
| Aroclor-1254-2 | (2) | 6.60 | 6.60 | 6.50 | 6.70 | 0.00 |
| Aroclor-1254-3 | (3) | 6.97 | 6.97 | 6.87 | 7.07 | 0.00 |
| Aroclor-1254-4 | (4) | 7.25 | 7.26 | 7.16 | 7.36 | 0.01 |
| Aroclor-1254-5 | (5) | 7.67 | 7.68 | 7.58 | 7.78 | 0.01 |
| Tetrachloro-m-xylene | | 4.37 | 4.37 | 4.27 | 4.47 | 0.00 |
| Decachlorobiphenyl | | 10.08 | 10.09 | 9.99 | 10.19 | 0.01 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

## CALIBRATION VERIFICATION SUMMARY

**Contract:** ENTA05

**Lab Code:** CHEM **Case No.:** P4799 **SAS No.:** P4799 **SDG NO.:** P4799

**Continuing Calib Date:** 11/13/2024 **Initial Calibration Date(s):** 11/11/2024 11/12/2024

**Continuing Calib Time:** 23:54 **Initial Calibration Time(s):** 17:30 00:52

**GC Column:** ZB-MR2 **ID:** 0.32 **(mm)**

| COMPOUND | | CCAL RT | AVG RT | RT WINDOW FROM | RT WINDOW TO | DIFF RT |
|---|---|---|---|---|---|---|
| Aroclor-1254-1 | (1) | 5.52 | 5.52 | 5.42 | 5.62 | 0.00 |
| Aroclor-1254-2 | (2) | 5.67 | 5.67 | 5.57 | 5.77 | 0.00 |
| Aroclor-1254-3 | (3) | 6.07 | 6.07 | 5.97 | 6.17 | 0.00 |
| Aroclor-1254-4 | (4) | 6.30 | 6.30 | 6.20 | 6.40 | 0.00 |
| Aroclor-1254-5 | (5) | 6.72 | 6.72 | 6.62 | 6.82 | 0.00 |
| Tetrachloro-m-xylene | | 3.64 | 3.64 | 3.54 | 3.74 | 0.00 |
| Decachlorobiphenyl | | 8.63 | 8.64 | 8.54 | 8.74 | 0.01 |