

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL01 Date Analyzed: 11/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO107877.D Time Analyzed: 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.528	5.427	5.627	518.650	500.000	3.7
Aroclor-1016-2	5.550	5.449	5.649	513.450	500.000	2.7
Aroclor-1016-3	5.613	5.512	5.712	506.800	500.000	1.4
Aroclor-1016-4	5.710	5.609	5.809	517.960	500.000	3.6
Aroclor-1016-5	6.007	5.906	6.106	524.520	500.000	4.9
Aroclor-1260-1	7.138	7.037	7.237	524.620	500.000	4.9
Aroclor-1260-2	7.396	7.295	7.495	532.230	500.000	6.4
Aroclor-1260-3	7.760	7.658	7.858	512.650	500.000	2.5
Aroclor-1260-4	7.984	7.883	8.083	505.640	500.000	1.1
Aroclor-1260-5	8.300	8.198	8.398	528.360	500.000	5.7
Decachlorobiphenyl	10.094	9.991	10.191	53.090	50.000	6.2
Tetrachloro-m-xylene	4.374	4.272	4.472	52.160	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL01 Date Analyzed: 11/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO107877.D Time Analyzed: 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.722	4.623	4.823	527.870	500.000	5.6
Aroclor-1016-2	4.741	4.642	4.842	536.210	500.000	7.2
Aroclor-1016-3	4.917	4.818	5.018	538.710	500.000	7.7
Aroclor-1016-4	4.959	4.859	5.059	537.080	500.000	7.4
Aroclor-1016-5	5.172	5.073	5.273	526.580	500.000	5.3
Aroclor-1260-1	6.203	6.104	6.304	534.790	500.000	7.0
Aroclor-1260-2	6.391	6.292	6.492	529.670	500.000	5.9
Aroclor-1260-3	6.545	6.445	6.645	533.590	500.000	6.7
Aroclor-1260-4	7.016	6.916	7.116	540.510	500.000	8.1
Aroclor-1260-5	7.257	7.158	7.358	544.630	500.000	8.9
Decachlorobiphenyl	8.638	8.540	8.740	54.260	50.000	8.5
Tetrachloro-m-xylene	3.641	3.541	3.741	53.740	50.000	7.5