

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL03 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107907.D Time Analyzed: 01:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.427	5.627	544.450	500.000	8.9
Aroclor-1016-2	5.548	5.449	5.649	535.430	500.000	7.1
Aroclor-1016-3	5.610	5.512	5.712	527.310	500.000	5.5
Aroclor-1016-4	5.708	5.609	5.809	528.920	500.000	5.8
Aroclor-1016-5	6.004	5.906	6.106	510.870	500.000	2.2
Aroclor-1260-1	7.135	7.037	7.237	494.230	500.000	-1.2
Aroclor-1260-2	7.392	7.295	7.495	505.970	500.000	1.2
Aroclor-1260-3	7.755	7.658	7.858	471.270	500.000	-5.7
Aroclor-1260-4	7.980	7.883	8.083	443.440	500.000	-11.3
Aroclor-1260-5	8.294	8.198	8.398	462.420	500.000	-7.5
Decachlorobiphenyl	10.083	9.991	10.191	44.820	50.000	-10.4
Tetrachloro-m-xylene	4.372	4.272	4.472	54.480	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL03 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107907.D Time Analyzed: 01:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.623	4.823	543.590	500.000	8.7
Aroclor-1016-2	4.740	4.642	4.842	548.190	500.000	9.6
Aroclor-1016-3	4.916	4.818	5.018	549.660	500.000	9.9
Aroclor-1016-4	4.958	4.859	5.059	532.550	500.000	6.5
Aroclor-1016-5	5.171	5.073	5.273	537.280	500.000	7.5
Aroclor-1260-1	6.202	6.104	6.304	565.390	500.000	13.1
Aroclor-1260-2	6.389	6.292	6.492	555.580	500.000	11.1
Aroclor-1260-3	6.543	6.445	6.645	557.950	500.000	11.6
Aroclor-1260-4	7.013	6.916	7.116	561.830	500.000	12.4
Aroclor-1260-5	7.254	7.158	7.358	558.020	500.000	11.6
Decachlorobiphenyl	8.635	8.540	8.740	47.920	50.000	-4.2
Tetrachloro-m-xylene	3.640	3.541	3.741	55.040	50.000	10.1