

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL05 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107926.D Time Analyzed: 13:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.427	5.627	484.940	500.000	-3.0
Aroclor-1016-2	5.547	5.449	5.649	474.270	500.000	-5.1
Aroclor-1016-3	5.610	5.512	5.712	474.750	500.000	-5.1
Aroclor-1016-4	5.707	5.609	5.809	485.980	500.000	-2.8
Aroclor-1016-5	6.003	5.906	6.106	480.500	500.000	-3.9
Aroclor-1260-1	7.133	7.037	7.237	454.970	500.000	-9.0
Aroclor-1260-2	7.390	7.295	7.495	463.840	500.000	-7.2
Aroclor-1260-3	7.753	7.658	7.858	423.270	500.000	-15.3
Aroclor-1260-4	7.979	7.883	8.083	418.820	500.000	-16.2
Aroclor-1260-5	8.293	8.198	8.398	434.950	500.000	-13.0
Decachlorobiphenyl	10.082	9.991	10.191	43.550	50.000	-12.9
Tetrachloro-m-xylene	4.372	4.272	4.472	48.290	50.000	-3.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL05 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107926.D Time Analyzed: 13:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.623	4.823	493.240	500.000	-1.4
Aroclor-1016-2	4.739	4.642	4.842	493.160	500.000	-1.4
Aroclor-1016-3	4.915	4.818	5.018	498.390	500.000	-0.3
Aroclor-1016-4	4.956	4.859	5.059	501.320	500.000	0.3
Aroclor-1016-5	5.170	5.073	5.273	479.450	500.000	-4.1
Aroclor-1260-1	6.200	6.104	6.304	494.350	500.000	-1.1
Aroclor-1260-2	6.387	6.292	6.492	490.700	500.000	-1.9
Aroclor-1260-3	6.541	6.445	6.645	493.330	500.000	-1.3
Aroclor-1260-4	7.012	6.916	7.116	495.370	500.000	-0.9
Aroclor-1260-5	7.253	7.158	7.358	496.330	500.000	-0.7
Decachlorobiphenyl	8.634	8.540	8.740	47.340	50.000	-5.3
Tetrachloro-m-xylene	3.639	3.541	3.741	48.870	50.000	-2.3