

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** P4799 **SAS No.:** P4799 **SDG NO.:** P4799

Continuing Calib Date: 11/14/2024 **Initial Calibration Date(s):** 11/13/2024 11/13/2024

Continuing Calib Time: 20:40 **Initial Calibration Time(s):** 11:44 13:23

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.24	7.24	7.14	7.34	0.00
2,4-D	8.37	8.37	8.27	8.47	0.00
2,4,5-TP(Silvex)	9.25	9.25	9.15	9.35	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** P4799 **SAS No.:** P4799 **SDG NO.:** P4799

Continuing Calib Date: 11/14/2024 **Initial Calibration Date(s):** 11/13/2024 11/13/2024

Continuing Calib Time: 20:40 **Initial Calibration Time(s):** 11:44 13:23

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.74	7.74	7.64	7.84	0.00
2,4-D	8.99	8.99	8.89	9.09	0.00
2,4,5-TP(Silvex)	9.89	9.90	9.80	10.00	0.01