

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: HSTDCCC750 Data File : PS028516.D Time Analyzed: 23:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.249	9.150	9.350	734.620	712.500	3.1
2,4-D	8.368	8.270	8.470	716.820	705.000	1.7
2,4-DCAA	7.240	7.141	7.341	761.470	750.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: HSTDCCC750 Data File : PS028516.D Time Analyzed: 23:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.894	9.796	9.996	750.590	712.500	5.3
2,4-D	8.987	8.890	9.090	730.080	705.000	3.6
2,4-DCAA	7.740	7.642	7.842	775.080	750.000	3.3