

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL02 Date Analyzed: 11/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS028528.D Time Analyzed: 04:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.249	9.150	9.350	747.710	712.500	4.9
2,4-D	8.369	8.270	8.470	726.840	705.000	3.1
2,4-DCAA	7.240	7.141	7.341	754.230	750.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL02 Date Analyzed: 11/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS028528.D Time Analyzed: 04:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.894	9.796	9.996	765.010	712.500	7.4
2,4-D	8.988	8.890	9.090	746.550	705.000	5.9
2,4-DCAA	7.740	7.642	7.842	790.710	750.000	5.4