

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111324\
 Data File : PS028516.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2024 23:13
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 143 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 23:27:10 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 13 13:41:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.240	7.740	1946.2E6	1001.6E6	761.468	775.076
Target Compounds							
1) T	Dalapon	2.638	2.698	2348.8E6	1676.7E6	699.145	692.860
2) T	3,5-DICHL...	6.410	6.693	2667.0E6	1402.7E6	714.773	725.909
3) T	4-Nitroph...	7.038	7.266	1181.2E6	601.2E6	695.454	696.316
5) T	DICAMBA	7.428	7.940	7961.6E6	4243.2E6	727.621	743.898
6) T	MCPD	7.612	8.043	525.4E6	311.4E6	74.683	72.403
7) T	MCPA	7.762	8.288	736.4E6	437.3E6	72.161	70.599
8) T	DICHLORPROP	8.138	8.657	2116.0E6	1067.2E6	716.970	730.695
9) T	2,4-D	8.368	8.987	2291.6E6	1143.0E6	716.821	730.082
10) T	Pentachlo...	8.668	9.516	31947.9E6	15192.5E6	765.475	757.908
11) T	2,4,5-TP ...	9.249	9.894	12760.8E6	6333.7E6	734.616	750.590
12) T	2,4,5-T	9.541	10.315	13040.7E6	6266.5E6	733.670	747.588
13) T	2,4-DB	10.117	10.882	1965.0E6	769.0E6	720.921	733.963
14) T	DINOSEB	11.332	11.261	10304.7E6	4036.0E6	722.745	730.034
15) T	Picloram	11.137	12.357	19994.8E6	8160.7E6	710.050	720.969
16) T	DCPA	11.625	12.307	19552.1E6	7371.9E6	743.049	760.111

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111324\
Data File : PS028516.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2024 23:13
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 143 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 23:27:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
Quant Title : 8080.M
QLast Update : Wed Nov 13 13:41:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

