

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4799

Client: ENTACT

Analytical Method: 8151A

DataFile : PS028522.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: P4799-03MSD	WC-TA2-02-CMSD											
	2,4-D	50	0	53.3	ug/L	107		1		70 (65)	130 (135)	20 (20)
	2,4,5-TP(Silvex)	50	0	71.9	ug/L	144	*	3		70 (62)	130 (139)	20 (20)