

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** P4799

**Client:** ENTACT

**Analytical Method:** **8151A** **Datafile :** PS028518.D

| Lab Sample ID | Parameter        | Spike | Result | Units | Rec | RPD | Qual | RPD  |         | Limits    |     |
|---------------|------------------|-------|--------|-------|-----|-----|------|------|---------|-----------|-----|
|               |                  |       |        |       |     |     |      | Qual | Low     | High      | RPD |
| PB164959BS    | 2,4-D            | 5     | 5.30   | ug/L  | 106 |     |      |      | 70 (83) | 130 (130) |     |
|               | 2,4,5-TP(Silvex) | 5     | 5.20   | ug/L  | 104 |     |      |      | 70 (78) | 130 (127) |     |