

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093039.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.052	8.954	9.154	44.920	50.000	-10.2
Endrin	6.574	6.474	6.674	43.880	50.000	-12.2
gamma-BHC (Lindane)	4.326	4.228	4.428	48.610	50.000	-2.8
Heptachlor	4.915	4.817	5.017	46.490	50.000	-7.0
Heptachlor epoxide	5.682	5.584	5.784	47.320	50.000	-5.4
Methoxychlor	7.498	7.399	7.599	44.300	50.000	-11.4
Tetrachloro-m-xylene	3.538	3.440	3.640	49.310	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093039.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.913	7.816	8.016	47.140	50.000	-5.7
Endrin	5.638	5.541	5.741	51.820	50.000	3.6
gamma-BHC (Lindane)	3.608	3.511	3.711	52.820	50.000	5.6
Heptachlor	3.947	3.849	4.049	51.830	50.000	3.7
Heptachlor epoxide	4.728	4.632	4.832	52.680	50.000	5.4
Methoxychlor	6.612	6.515	6.715	49.140	50.000	-1.7
Tetrachloro-m-xylene	2.775	2.678	2.878	51.640	50.000	3.3