

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093051.D Time Analyzed: 19:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.053	8.954	9.154	44.800	50.000	-10.4
Endrin	6.573	6.474	6.674	43.110	50.000	-13.8
gamma-BHC (Lindane)	4.327	4.228	4.428	47.910	50.000	-4.2
Heptachlor	4.915	4.817	5.017	45.940	50.000	-8.1
Heptachlor epoxide	5.681	5.584	5.784	47.660	50.000	-4.7
Methoxychlor	7.498	7.399	7.599	43.930	50.000	-12.1
Tetrachloro-m-xylene	3.539	3.440	3.640	48.420	50.000	-3.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093051.D Time Analyzed: 19:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.912	7.816	8.016	50.100	50.000	0.2
Endrin	5.639	5.541	5.741	51.360	50.000	2.7
gamma-BHC (Lindane)	3.609	3.511	3.711	52.250	50.000	4.5
Heptachlor	3.947	3.849	4.049	51.030	50.000	2.1
Heptachlor epoxide	4.728	4.632	4.832	52.530	50.000	5.1
Methoxychlor	6.612	6.515	6.715	49.080	50.000	-1.8
Tetrachloro-m-xylene	2.776	2.678	2.878	51.220	50.000	2.4