

### Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4799

Client: ENTACT

Analytical Method: 8081B

DataFile : PL093045.D

| Lab Sample ID:           | Parameter             | Spike | Sample<br>Result | Result | Units | Rec | Rec<br>Qual | RPD | RPD<br>Qual | Low     | Limits<br>High | RPD     |
|--------------------------|-----------------------|-------|------------------|--------|-------|-----|-------------|-----|-------------|---------|----------------|---------|
| <b>Client Sample ID:</b> | <b>WC-TA2-02-CMSD</b> |       |                  |        |       |     |             |     |             |         |                |         |
| P4799-03MSD              | gamma-BHC (Lindane)   | 5     | 0                | 5.80   | ug/L  | 116 |             | 0   |             | 30 (60) | 150 (152)      | 20 (20) |
|                          | Heptachlor            | 5     | 0                | 9.20   | ug/L  | 184 | *           | 0   |             | 30 (56) | 150 (147)      | 20 (20) |
|                          | Heptachlor epoxide    | 5     | 0                | 6.20   | ug/L  | 124 |             | 0   |             | 30 (77) | 150 (143)      | 20 (20) |
|                          | Endrin                | 5     | 0                | 6.50   | ug/L  | 130 |             | 0   |             | 30 (76) | 150 (144)      | 20 (20) |
|                          | Methoxychlor          | 5     | 0                | 6.10   | ug/L  | 122 |             | 2   |             | 30 (70) | 150 (142)      | 20 (20) |