

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4799

Client: ENTACT

Analytical Method: **8081B** Datafile : PL093041.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164960BS	gamma-BHC (Lindane)	0.5	0.55	ug/L	111				40 (82)	140 (129)	
	Heptachlor	0.5	0.58	ug/L	117				40 (79)	140 (127)	
	Heptachlor epoxide	0.5	0.58	ug/L	116				40 (81)	140 (124)	
	Endrin	0.5	0.61	ug/L	121				40 (81)	140 (128)	
	Methoxychlor	0.5	0.57	ug/L	114				40 (78)	140 (108)	