

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG No.: P4799

Instrument ID: MSVOA_N Calibration Date/Time: 11/12/2024 12:25

Lab File ID: VN084789.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.618	0.674		9.06	20
1,1-Dichloroethene	0.569	0.567		-0.35	20
2-Butanone	0.337	0.404		19.88	20
Carbon Tetrachloride	0.525	0.578		10.1	20
Chloroform	1.121	1.175		4.82	20
Benzene	1.507	1.565		3.85	20
1,2-Dichloroethane	0.488	0.537		10.04	20
Trichloroethene	0.348	0.363		4.31	20
Tetrachloroethene	0.333	0.364		9.31	20
Chlorobenzene	1.119	1.182	0.3	5.63	20
1,2-Dichloroethane-d4	0.722	0.706		-2.22	20
Dibromofluoromethane	0.338	0.350		3.55	20
Toluene-d8	1.247	1.248		0.08	20
4-Bromofluorobenzene	0.466	0.478		2.58	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.