

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4799

Client: ENTACT

Analytical Method: SW8260-Low

Datafile : VN084792.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1112WBS01	Vinyl chloride	20	19.0	ug/L	95			70 (65)	130 (117)	
	1,1-Dichloroethene	20	16.1	ug/L	81			70 (74)	130 (110)	
	2-Butanone	100	93.8	ug/L	94			40 (65)	160 (122)	
	Carbon Tetrachloride	20	18.0	ug/L	90			70 (77)	130 (113)	
	Chloroform	20	16.9	ug/L	85			70 (79)	130 (113)	
	Benzene	20	17.2	ug/L	86			70 (82)	130 (109)	
	1,2-Dichloroethane	20	17.7	ug/L	89			70 (80)	130 (115)	
	Trichloroethene	20	16.9	ug/L	85			70 (77)	130 (113)	
	Tetrachloroethene	20	18.6	ug/L	93			70 (67)	130 (123)	
	Chlorobenzene	20	17.7	ug/L	89			70 (82)	130 (109)	