284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4799

Client: ENTACT

Analytical Method: SW8260-Low **Datafile :** VN084822.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	Limits RPD
VN1113WBS01	Vinyl chloride	20	20.1	ug/L	101			70 (65)	130 (117)	
	1,1-Dichloroethene	20	17.5	ug/L	88			70 (74)	130 (110)	
	2-Butanone	100	91.6	ug/L	92			40 (65)	160 (122)	
	Carbon Tetrachloride	20	20.1	ug/L	101			70 (77)	130 (113)	
	Chloroform	20	18.9	ug/L	95			70 (79)	130 (113)	
	Benzene	20	18.5	ug/L	93			70 (82)	130 (109)	
	1,2-Dichloroethane	20	19.0	ug/L	95			70 (80)	130 (115)	
	Trichloroethene	20	19.2	ug/L	96			70 (77)	130 (113)	
	Tetrachloroethene	20	20.6	ug/L	103			70 (67)	130 (123)	
	Chlorobenzene	20	18.9	ug/L	95			70 (82)	130 (109)	

( ) = LABORATORY INHOUSE LIMIT