

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799
Lab File ID: VN084789.D Date Analyzed: 11/12/2024
Instrument ID: MSVOA_N Time Analyzed: 12:25
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	154488	8.22	250919	9.10	222019	11.87
UPPER LIMIT	308976	8.724	501838	9.6	444038	12.365
LOWER LIMIT	77244	7.724	125460	8.6	111010	11.365
EPA SAMPLE NO.						
WC-TA2-02-G	173752	8.22	312071	9.10	271773	11.87
WC-TA2-02-G	173752	8.22	312071	9.10	271773	11.87
WC-TA2-03-G	170276	8.22	297743	9.10	265314	11.87
WC-TA2-03-G	170276	8.22	297743	9.10	265314	11.87
WC-TA1-01-G	169069	8.22	287640	9.10	266403	11.87
WC-TA1-01-G	169069	8.22	287640	9.10	266403	11.87
WC-TA1-02-G	178622	8.22	301620	9.09	283346	11.87
WC-TA1-02-G	178622	8.22	301620	9.09	283346	11.87
VN1112WBL01	180518	8.22	306816	9.10	261220	11.87
VN1112WBL01	180518	8.22	306816	9.10	261220	11.87
VN1112WBS01	181136	8.22	298646	9.10	255962	11.87
VN1112WBS01	181136	8.22	298646	9.10	255962	11.87
VN1112WBSD01	169278	8.22	279342	9.10	250471	11.87
VN1112WBSD01	169278	8.22	279342	9.10	250471	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.