

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P4799 SAS No.: P4799 SDG NO.: P4799
 Lab File ID: VN084789.D Date Analyzed: 11/12/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	111546	13.794				
UPPER LIMIT	223092	14.294				
LOWER LIMIT	55773	13.294				
EPA SAMPLE NO.						
WC-TA2-02-G	120380	13.79				
WC-TA2-02-G	120380	13.79				
WC-TA2-03-G	124378	13.79				
WC-TA2-03-G	124378	13.79				
WC-TA1-01-G	141406	13.79				
WC-TA1-01-G	141406	13.79				
WC-TA1-02-G	147697	13.79				
WC-TA1-02-G	147697	13.79				
VN1112WBL01	118783	13.79				
VN1112WBL01	118783	13.79				
VN1112WBS01	130240	13.79				
VN1112WBS01	130240	13.79				
VN1112WBSD01	131269	13.79				
VN1112WBSD01	131269	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.