

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51822 MA No.: SDG No.: MH2GN9
 Instrument ID: P7 Start Date : 11/18/2024
 Run Batch: LB133504 End Date : 11/18/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		165Ho	Q	209Bi	Q		Q		Q		Q
S0	0431	100		100							
S02	0438	100		100							
S03	0441	100		101							
S04	0444	103		103							
S05	0447	102		102							
S06	0449	100		98							
S07	0452	97		94							
S08	0455	95		84							
ICV001	0553	100		101							
ICB001	0557	99		99							
ICSA001	0600	99		94							
ICSAB001	0603	103		97							
CCV001	0606	104		95							
CCB001	0609	105		106							
PBS950	0625	104		104							
LCS950	0628	102		101							
ZZZZZZ	0631										
ZZZZZZ	0635										
MH2GN9	0638	104		103							
MH2GN9D	0641	105		104							
MH2GN9L	0644	101		102							
MH2GN9S	0647	108		106							
MH2GP0	0650	106		105							
MH2GP1	0653	107		105							
MH2GP2	0656	108		105							
MH2GP3	0659	109		106							
MH2GP4	0702	106		104							
MH2GP5	0705	107		104							
MH2GP6	0708	105		103							
MH2GP7	0711	102		99							
MH2GP8	0714	107		106							
MH2GP9	0717	108		105							
MH2GQ0	0720	108		106							
MH2GQ1	0723	108		106							
ZZZZZZ	0726										
CCV002	0729	105		96							
CCB002	0731	102		103							

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Lab Name:	Alliance Technical Group, LLC	Contract:	68HERH20D0011
Lab Code:	ACE	Case No.:	51822
		MA No.:	SDG No.: MH2GN9
Instrument ID:	P7	Start Date :	11/18/2024
Run Batch:	LB133504	End Date :	11/18/2024

[illegible]

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51822 MA No.: _____ SDG No.: MH2GN9
 Instrument ID: P7 Start Date : 11/18/2024
 Run Batch: LB133504 End Date : 11/18/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		209Bi	Q		Q		Q		Q		Q
S0	0431	100									
S02	0438	101									
S03	0441	100									
S04	0444	98									
S05	0447	100									
S06	0449	98									
S07	0452	96									
S08	0455	83									
ICV001	0553	101									
ICB001	0557	101									
ICSA001	0600	92									
ICSAB001	0603	95									
CCV001	0606	92									
CCB001	0609	106									
PBS950	0625	103									
LCS950	0628	103									
ZZZZZZ	0631										
ZZZZZZ	0635										
MH2GN9	0638	100									
MH2GN9D	0641	101									
MH2GN9L	0644	103									
MH2GN9S	0647	101									
MH2GP0	0650	101									
MH2GP1	0653	102									
MH2GP2	0656	96									
MH2GP3	0659	102									
MH2GP4	0702	101									
MH2GP5	0705	99									
MH2GP6	0708	98									
MH2GP7	0711	100									
MH2GP8	0714	102									
MH2GP9	0717	103									
MH2GQ0	0720	102									
MH2GQ1	0723	102									
ZZZZZZ	0726										
CCV002	0729	92									
CCB002	0731	103									

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Lab Name: Alliance Technical Group, LLC	Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51822	MA No.: SDG No.: MH2GN9
Instrument ID: P7	Start Date : 11/18/2024
Run Batch: LB133504	End Date : 11/18/2024

[illegible]