

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: SCIA01
 ProjectID: 124 Main St. Bloomingtondale
 Lab Code: CHEM Case No.: P4810 SAS No.: P4810 SDG No.: P4810

Calibration Sequence : FG111324		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
850	90369281	106317	FG014760.D
340	37189473	109381	FG014761.D
170	21164070	124495	FG014762.D
85	9606047	113012	FG014763.D
1700	168427130	99075	FG014764.D
AVG RF : 110456		% RSD : 8.486	AVG RT : 15.0442