

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: SCIA01
ProjectID: 124 Main St. Bloomingtondale
Lab Code: CHEM Case No.: P4810 SAS No.: P4810 SDG No.: P4810
DataFile: FG014799.D Analyst Name: YP\AJ Analyst Date: 11-15-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
850	96219103	113199	110456	2.483