284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

Client: Sciacca General Contractors, LLC **SDG No.:** P4810

Project: 124 Main St. Bloomingdale **Instrument ID:** FID_G

GC Column: RXI-1MS **ID:** 0.18 **(mm)**

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES, AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION **15.0442**

EPA SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
PIBLK01	I.BLK01	15 Nov 2024 09:56	FG014790.D	15.041	
50 PPM TRPH STD	50 PPM TRPH STD	15 Nov 2024 10:24	FG014791.D	15.045	
PB164989BL	PB164989BL	15 Nov 2024 11:24	FG014793.D	15.041	
PB164989BS	PB164989BS	15 Nov 2024 11:53	FG014794.D	15.041	
WASTE	P4810-01	15 Nov 2024 12:21	FG014795.D	15.042	
WASTEMS	P4810-01MS	15 Nov 2024 12:49	FG014796.D	15.042	
WASTEMSD	P4810-01MSD	15 Nov 2024 13:18	FG014797.D	15.043	
PIBLK02	I.BLK02	15 Nov 2024 13:46	FG014798.D	15.043	
50 PPM TRPH STD	50 PPM TRPH STD	15 Nov 2024 14:14	FG014799.D	15.045	

 # Column used to flag RT values with an * values outside of QC limits

QC Limits	Lower Limit	Upper Limits
(± 0.10 minutes)	14.9442	15.1442