

Pratik Patel

From: Bett, Daisy <Daisy.Bett@gdit.com>
Sent: Tuesday, November 12, 2024 4:36 PM
To: Deepak Parmar; Sohil Jodhani; Mohammad Ahmed
Cc: Leung.christina@epa.gov; Feranda, Jennifer; Brandon-Bazile, Kim; Bauer, Heather E; Johnson, Matthew
Subject: Region 02 | Case 51793 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC | FINAL
Attachments: SKM_95824111209592.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,


Issue: The attached COC designates two samples for laboratory QC but only one QC sample is required for each 20 sample SDG. The laboratory would like to confirm that they can proceed with sample BHK70 only for QC analysis and keep sample BHK71 as a regular sample.

Resolution: Per Region 2, the laboratory should proceed with sample BHK71 as the Laboratory QC sample and analyze sample BHK70 as a regular sample. Please note the issue in the SDG Narrative and proceed with the analysis of the samples.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Thank you,
Daisy Bett
Research Analyst Associate
GDIT Federal Civilian Division
EPA Region 2&3 CLP QSS Coordinator
Under contract to the EPA

T: 571.454.0186
daisy.bett@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



Leave alert: Nov 29th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>
Sent: Tuesday, November 12, 2024 2:46 PM
To: Bett, Daisy <Daisy.Bett@gdit.com>
Cc: Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>
Subject: FW: Region 02 | Case 51793 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Daisy,

Sampler's response is below.

Regards,

Christina Leung
Regional Sample Control Center (RSCC)
USEPA Region 2
LSASD-HWSB-HWSS
732-906-6995
Leung.christina@epa.gov
Updated CLPSS Address: <https://clpss.epa.gov/uaa/login>

From: Tanwar, Simriti <Simriti.Tanwar@aeacom.com>
Sent: Tuesday, November 12, 2024 2:32 PM
To: Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Raduma, Modi <modi.raduma@aeacom.com>; Purdy, Richard <richard.purdy@aeacom.com>; Ketu, Rupika <Ketu.Rupika@epa.gov>
Cc: Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Toomey, Courtney <courtney.toomey@aeacom.com>; Zeidan, Mousa <Mousa.Zeidan@aeacom.com>
Subject: RE: Region 02 | Case 51793 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Hi Christina,

Modi is out of office currently and Courtney Toomey (copied on this email) is covering for her. Please keep BHK70 as the regular sample and BHK71 as the QC sample. Please let us know if there are any further questions.

Thanks
Simriti Tanwar, PMP

Senior Project Manager, Environment
O. 732-564-3600
D. 732-564-3645
C. 732-208-3599
simriti.tanwar@aeacom.com

AECOM

30 Knightsbridge Road
Suite 520
Piscataway, New Jersey 08854

From: Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>
Sent: Tuesday, November 12, 2024 1:39 PM
To: Raduma, Modi <Modi.Raduma@aeacom.com>; Purdy, Richard <Richard.Purdy@aeacom.com>; Tanwar, Simriti <Simriti.Tanwar@aeacom.com>; Ketu, Rupika <Ketu.Rupika@epa.gov>
Cc: Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>
Subject: FW: Region 02 | Case 51793 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

Hi Modi,

Could you please advise?

Regards,

Christina Leung

Regional Sample Control Center (RSCC)
USEPA Region 2
LSASD-HWSB-HWSS
732-906-6995
Leung.christina@epa.gov
Updated CLPSS Address: <https://clpss.epa.gov/uaa/login>

From: Bett, Daisy <Daisy.Bett@gdit.com>
Sent: Tuesday, November 12, 2024 12:55 PM
To: Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>
Cc: Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>
Subject: Region 02 | Case 51793 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon,

Issue: The attached COC designates two samples for laboratory QC but only one QC sample is required for each 20 sample SDG. The laboratory would like to confirm that they can proceed with sample BHK70 only for QC analysis and keep sample BHK71 as a regular sample.

Thank you,
Daisy Bett
Research Analyst Associate
GDIT Federal Civilian Division

EPA Region 2&3 CLP QSS Coordinator

Under contract to the EPA


T: 571.454.0186

daisy.bett@gdit.com

15036 Conference Center Drive

Chantilly, VA 20151

www.gdit.com



GENERAL DYNAMICS
Information Technology

Leave alert: Nov 29th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DEEPAK@CHEMTECH.NET <DEEPAK@CHEMTECH.NET>

Sent: Tuesday, November 12, 2024 10:19 AM

To: Bett, Daisy <Daisy.Bett@gdit.com>

Cc: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>

Subject: Region 2 | Case 51793 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

Good morning,

Issue 1 : Only one QC sample required for each 20 sample SDG, however two sample mentioned on COC for Lab QC. Therefor lab will use only sample BHK70 as QC other sample will just analysis as regular sample.

Please see attachment for your reference.

Thanks & Regards,



Deepak Parmar

QA/QC

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3154

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com

