

Laboratory Control Sample Summary

Client: BAPS North Bergen

SDG No.: P4822

Project: BAPS North Bergen

Run No.: LB133487

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165016BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	11/15/2024