

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: BAPS North Bergen **SDG No.:** P4822
Contract: BAPS03 **Lab Code:** CHEM **Case No.:** P4822 **SAS No.:** P4822

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164935BS							
Aluminum	mg/Kg	85.5	78.7		92	80 - 120	P
Antimony	mg/Kg	34.2	30.5		89	80 - 120	P
Arsenic	mg/Kg	34.2	31.1		91	80 - 120	P
Barium	mg/Kg	8.5	7.95		94	80 - 120	P
Beryllium	mg/Kg	8.5	8.23		97	80 - 120	P
Cadmium	mg/Kg	8.5	7.58		89	80 - 120	P
Calcium	mg/Kg	42.7	41.7	J	98	80 - 120	P
Chromium	mg/Kg	17.1	15.9		93	80 - 120	P
Cobalt	mg/Kg	8.5	7.72		91	80 - 120	P
Copper	mg/Kg	12.8	12.2		95	80 - 120	P
Iron	mg/Kg	130	117		90	80 - 120	P
Lead	mg/Kg	42.7	38.6		90	80 - 120	P
Magnesium	mg/Kg	85.5	78.7	J	92	80 - 120	P
Manganese	mg/Kg	8.5	8.22		97	80 - 120	P
Nickel	mg/Kg	21.4	19.4		91	80 - 120	P
Potassium	mg/Kg	430	375		87	80 - 120	P
Selenium	mg/Kg	85.5	75.6		88	80 - 120	P
Silver	mg/Kg	3.2	2.94		92	80 - 120	P
Sodium	mg/Kg	130	105		81	80 - 120	P
Thallium	mg/Kg	85.5	79.5		93	80 - 120	P
Vanadium	mg/Kg	12.8	12.0		94	80 - 120	P
Zinc	mg/Kg	8.5	8.21		97	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: BAPS North Bergen

SDG No.: P4822

Contract: BAPS03

Lab Code: CHEM

Case No.: P4822

SAS No.: P4822

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164947BS Mercury	mg/Kg	0.27	0.24		87	80 - 120	CV