

Continuing Calibration Report for SequenceID : FC111324AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067745.D

Aliphatic C9-C12	7960875.000	60.000	3.169	6.447	132681.250	135469.641	2.058
Aliphatic C12-C16	5183286.000	40.000	6.448	9.836	129582.150	131681.730	1.594
Aliphatic C16-C21	7536116.000	60.000	9.837	13.192	125601.933	127235.092	1.284
Aliphatic C21-C28	9658304.000	80.000	13.193	16.845	120728.800	120090.398	-0.532
Aliphatic C28-C40	12823575.000	120.000	16.846	21.681	106863.125	101093.195	-5.708
Aliphatic EPH	43162156.000	360.000	3.169	21.681	119894.878	118799.912	-0.922