

Continuing Calibration Report for SequenceID : FD111324AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048715.D

Aromatic C10-C12	7611262.000	40.000	4.089	5.807	190281.550	176202.694	-7.990
Aromatic C12-C16	11365912.000	60.000	5.808	8.412	189431.867	177273.493	-6.859
Aromatic C16-C21	13442306.000	80.000	8.413	12.673	168028.825	163832.550	-2.561
Aromatic C21-C36	30749875.000	180.000	12.674	18.078	170832.639	152713.153	-11.865
Aromatic EPH	63169355.000	360.000	4.089	18.078	175470.431	161887.469	-8.390