

Continuing Calibration Report for SequenceID : FD111324AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048726.D

Aromatic C10-C12	6232410.000	40.000	4.089	5.807	155810.250	176202.694	11.573
Aromatic C12-C16	9308929.000	60.000	5.808	8.412	155148.817	177273.493	12.481
Aromatic C16-C21	11595850.000	80.000	8.413	12.673	144948.125	163832.550	11.527
Aromatic C21-C36	25181205.000	180.000	12.674	18.078	139895.583	152713.153	8.393
Aromatic EPH	52318394.000	360.000	4.089	18.078	145328.872	161887.469	10.228