

Continuing Calibration Report for SequenceID : FF111324AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015011.D

Aromatic EPH	50600951.000	360.000	4.508	18.815	140558.197	143365.296	1.958
Aromatic C10-C12	6223435.000	40.000	4.508	6.372	155585.875	154569.237	-0.658
Aromatic C12-C16	9170599.000	60.000	6.373	9.063	152843.317	153560.844	0.467
Aromatic C16-C21	11198491.000	80.000	9.064	13.374	139981.138	142096.237	1.488
Aromatic C21-C36	24008426.000	180.000	13.375	18.815	133380.144	138041.042	3.376