

Continuing Calibration Report for SequenceID : FF111324AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015023.D

Aromatic EPH	50197626.000	360.000	4.508	18.815	139437.850	143365.296	2.739
Aromatic C10-C12	6068161.000	40.000	4.508	6.372	151704.025	154569.237	1.854
Aromatic C12-C16	9060125.000	60.000	6.373	9.063	151002.083	153560.844	1.666
Aromatic C16-C21	11097245.000	80.000	9.064	13.374	138715.563	142096.237	2.379
Aromatic C21-C36	23972095.000	180.000	13.375	18.815	133178.306	138041.042	3.523