

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** BAPS North Bergen
Lab Code: CHEM **Cas No:** P4822 **SAS No :** P4822 **SDG No:** P4822
Sample No : P4768-07MSD **Datafile:** FC067741.D
Client ID : B-6-3MSD

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS QC Limit Of RPD
Aliphatic C9-C12	10.5	0.42	7.56	68		3.6 (40-140) 25
Aliphatic C12-C16	7	0.53	7.24	96		3.07 (40-140) 25
Aliphatic C16-C21	10.5	0.40	11.4	105		1.89 (40-140) 25
Aliphatic C21-C28	13.9	0	15.5	111		1.79 (40-140) 25
Aliphatic C28-C40	20.9	6.13	37.1	148	*	2.74 (40-140) 25