

### CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

Continuing Calib Date: 11/13/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 14:26 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1248-1       | (1) | 5.52       | 5.53      | 5.43      | 5.63  | 0.01       |
| Aroclor-1248-2       | (2) | 5.80       | 5.80      | 5.70      | 5.90  | 0.00       |
| Aroclor-1248-3       | (3) | 6.00       | 6.00      | 5.90      | 6.10  | 0.00       |
| Aroclor-1248-4       | (4) | 6.40       | 6.41      | 6.31      | 6.51  | 0.01       |
| Aroclor-1248-5       | (5) | 6.44       | 6.45      | 6.35      | 6.55  | 0.01       |
| Tetrachloro-m-xylene |     | 4.37       | 4.37      | 4.27      | 4.47  | 0.00       |
| Decachlorobiphenyl   |     | 10.08      | 10.09     | 9.99      | 10.19 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

Continuing Calib Date: 11/13/2024 Initial Calibration Date(s): 11/11/2024 11/12/2024

Continuing Calib Time: 14:26 Initial Calibration Time(s): 17:30 00:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.72       | 4.72      | 4.62      | 4.82 | 0.00       |
| Aroclor-1248-2       | (2) | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1248-3       | (3) | 5.00       | 5.00      | 4.90      | 5.10 | 0.00       |
| Aroclor-1248-4       | (4) | 5.17       | 5.17      | 5.07      | 5.27 | 0.00       |
| Aroclor-1248-5       | (5) | 5.56       | 5.56      | 5.46      | 5.66 | 0.00       |
| Tetrachloro-m-xylene |     | 3.64       | 3.64      | 3.54      | 3.74 | 0.00       |
| Decachlorobiphenyl   |     | 8.63       | 8.64      | 8.54      | 8.74 | 0.01       |