

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107913.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.527	5.427	5.627	493.760	500.000	-1.2
Aroclor-1016-2	5.549	5.449	5.649	496.500	500.000	-0.7
Aroclor-1016-3	5.612	5.512	5.712	484.690	500.000	-3.1
Aroclor-1016-4	5.709	5.609	5.809	498.610	500.000	-0.3
Aroclor-1016-5	6.005	5.906	6.106	504.740	500.000	0.9
Aroclor-1260-1	7.136	7.037	7.237	434.020	500.000	-13.2
Aroclor-1260-2	7.394	7.295	7.495	457.920	500.000	-8.4
Aroclor-1260-3	7.756	7.658	7.858	435.860	500.000	-12.8
Aroclor-1260-4	7.981	7.883	8.083	437.480	500.000	-12.5
Aroclor-1260-5	8.296	8.198	8.398	456.540	500.000	-8.7
Decachlorobiphenyl	10.085	9.991	10.191	43.480	50.000	-13.0
Tetrachloro-m-xylene	4.374	4.272	4.472	49.790	50.000	-0.4

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107913.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.623	4.823	499.190	500.000	-0.2
Aroclor-1016-2	4.740	4.642	4.842	507.120	500.000	1.4
Aroclor-1016-3	4.916	4.818	5.018	509.280	500.000	1.9
Aroclor-1016-4	4.957	4.859	5.059	513.790	500.000	2.8
Aroclor-1016-5	5.171	5.073	5.273	496.710	500.000	-0.7
Aroclor-1260-1	6.201	6.104	6.304	505.810	500.000	1.2
Aroclor-1260-2	6.389	6.292	6.492	504.840	500.000	1.0
Aroclor-1260-3	6.543	6.445	6.645	508.080	500.000	1.6
Aroclor-1260-4	7.014	6.916	7.116	514.400	500.000	2.9
Aroclor-1260-5	7.255	7.158	7.358	517.800	500.000	3.6
Decachlorobiphenyl	8.635	8.540	8.740	49.150	50.000	-1.7
Tetrachloro-m-xylene	3.640	3.541	3.741	50.890	50.000	1.8