

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL03 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107940.D Time Analyzed: 18:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.523	5.427	5.627	479.630	500.000	-4.1
Aroclor-1016-2	5.545	5.449	5.649	474.960	500.000	-5.0
Aroclor-1016-3	5.608	5.512	5.712	463.160	500.000	-7.4
Aroclor-1016-4	5.705	5.609	5.809	472.970	500.000	-5.4
Aroclor-1016-5	6.001	5.906	6.106	477.530	500.000	-4.5
Aroclor-1260-1	7.132	7.037	7.237	445.490	500.000	-10.9
Aroclor-1260-2	7.389	7.295	7.495	460.220	500.000	-8.0
Aroclor-1260-3	7.752	7.658	7.858	425.120	500.000	-15.0
Aroclor-1260-4	7.978	7.883	8.083	411.020	500.000	-17.8
Aroclor-1260-5	8.291	8.198	8.398	428.490	500.000	-14.3
Decachlorobiphenyl	10.079	9.991	10.191	42.360	50.000	-15.3
Tetrachloro-m-xylene	4.370	4.272	4.472	49.290	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/11/2024 11/11/2024

Client Sample No.: CCAL03 Date Analyzed: 11/13/2024

Lab Sample No.: AR1660CCC500 Data File : PO107940.D Time Analyzed: 18:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.623	4.823	505.060	500.000	1.0
Aroclor-1016-2	4.738	4.642	4.842	505.040	500.000	1.0
Aroclor-1016-3	4.914	4.818	5.018	509.000	500.000	1.8
Aroclor-1016-4	4.956	4.859	5.059	506.960	500.000	1.4
Aroclor-1016-5	5.169	5.073	5.273	482.000	500.000	-3.6
Aroclor-1260-1	6.200	6.104	6.304	508.250	500.000	1.7
Aroclor-1260-2	6.388	6.292	6.492	502.150	500.000	0.4
Aroclor-1260-3	6.541	6.445	6.645	503.790	500.000	0.8
Aroclor-1260-4	7.012	6.916	7.116	499.310	500.000	-0.1
Aroclor-1260-5	7.252	7.158	7.358	497.970	500.000	-0.4
Decachlorobiphenyl	8.632	8.540	8.740	45.040	50.000	-9.9
Tetrachloro-m-xylene	3.639	3.541	3.741	49.670	50.000	-0.7