

Surrogate Summary

SDG No.: **P4822**

Client: **BAPS North Bergen**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107842.D	PIBLK-PO107842.D	Tetrachloro-m-xylene	1	20	21.0	105		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.2	116		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.2	111		70 (60)	130 (140)
I.BLK-PO107916.D	PIBLK-PO107916.D	Tetrachloro-m-xylene	1	20	20.9	105		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.5	103		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.1	111		70 (60)	130 (140)
PB164938BL	PB164938BL	Tetrachloro-m-xylene	1	20	18.6	93		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	18.5	92		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.4	92		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.1	96		30 (32)	150 (175)
PB164938BS	PB164938BS	Tetrachloro-m-xylene	1	20	19.4	97		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.4	97		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.5	93		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.5	102		30 (32)	150 (175)
I.BLK-PO107930.D	PIBLK-PO107930.D	Tetrachloro-m-xylene	1	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.8	104		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.5	103		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.7	108		70 (60)	130 (140)
P4822-02	SOIL-1	Tetrachloro-m-xylene	1	20	20.5	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	18.7	94		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.0	95		30 (32)	150 (175)
P4822-04	SOIL-2	Tetrachloro-m-xylene	1	20	21.7	109		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.7	99		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.8	109		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.1	101		30 (32)	150 (175)
P4822-06	SOIL-3	Tetrachloro-m-xylene	1	20	17.5	87		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	15.6	78		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.4	87		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	15.4	77		30 (32)	150 (175)
P4822-06MS	SOIL-3MS	Tetrachloro-m-xylene	1	20	22.1	110		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.2	101		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.3	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.9	99		30 (32)	150 (175)
P4822-06MSD	SOIL-3MSD	Tetrachloro-m-xylene	1	20	21.2	106		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.8	99		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.1	96		30 (32)	150 (175)
P4822-08	SOIL-4	Tetrachloro-m-xylene	1	20	20.5	102		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: **P4822**

Client: **BAPS North Bergen**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P4822-08	SOIL-4	Decachlorobiphenyl	1	20	19.2	96		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.3	107		30 (32)	150 (144)
P4822-10	SOIL-5	Decachlorobiphenyl	2	20	18.5	92		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	18.7	93		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.0	85		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	16.3	81		30 (32)	150 (175)
P4822-12	SOIL-6	Tetrachloro-m-xylene	1	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.2	96		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	22.3	112		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.9	94		30 (32)	150 (175)
I.BLK-PO107944.D	PIBLK-PO107944.D	Tetrachloro-m-xylene	1	20	20.7	103		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.5	102		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.7	103		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.9	104		70 (60)	130 (140)