

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: BAPS03

Lab Code: CHEM **Case No.:** P4822 **SAS No.:** P4822 **SDG NO.:** P4822

Instrument ID: ECD_L **Calibration Date(s):** 10/28/2024 10/28/2024

Calibration Times: 14:43 15:36

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL092655.D</u>		CF 075 = <u>PL092656.D</u>			
CF 050 = <u>PL092657.D</u>		CF 025 = <u>PL092658.D</u>		CF 005 = <u>PL092659.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	1810270000	1801980000	1817980000	1967770000	2199620000	1919520000	9
4,4'-DDE	2254070000	2226390000	2230050000	2402470000	2732960000	2369190000	9
4,4'-DDT	1948940000	1933390000	1940720000	2113460000	2412250000	2069750000	10
Aldrin	2847480000	2802990000	2807290000	3047230000	3502400000	3001480000	10
alpha-BHC	3356530000	3271880000	3230400000	3397090000	3701950000	3391570000	5
alpha-Chlordane	2486640000	2469090000	2489960000	2701720000	3093310000	2648140000	10
beta-BHC	1343260000	1339050000	1368530000	1491800000	1684660000	1445460000	10
Decachlorobiphenyl	1738840000	1756630000	1819720000	1998760000	2308800000	1924550000	12
delta-BHC	3067660000	2980040000	2949010000	3119670000	3533000000	3129870000	8
Dieldrin	2486930000	2456960000	2476030000	2679410000	3078050000	2635480000	10
Endosulfan I	2320690000	2313950000	2348370000	2546170000	2977620000	2501360000	11
Endosulfan II	2159880000	2160930000	2205230000	2449960000	2943170000	2383830000	14
Endosulfan sulfate	1974140000	1979440000	2024740000	2239840000	2633550000	2170340000	13
Endrin	2111540000	2103000000	2121260000	2324460000	2723040000	2276660000	12
Endrin aldehyde	1712300000	1718280000	1758160000	1955910000	2245210000	1877970000	12
Endrin ketone	2253830000	2246660000	2273910000	2471090000	2868570000	2422810000	11
gamma-BHC (Lindane)	3198960000	3133030000	3104430000	3278360000	3583040000	3259560000	6
gamma-Chlordane	2496700000	2477960000	2491940000	2703470000	3133120000	2660640000	11
Heptachlor	2817300000	2795570000	2829220000	3064000000	3509480000	3003110000	10
Heptachlor epoxide	2536240000	2521530000	2566410000	2821600000	3361270000	2761410000	13
Methoxychlor	1040530000	1050870000	1078280000	1189160000	1341160000	1140000000	11
Tetrachloro-m-xylene	2319350000	2304070000	2328420000	2512350000	2786990000	2450240000	8

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

Instrument ID: ECD_L

Calibration Date(s): 10/28/2024 10/28/2024

Calibration Times: 14:43 15:36

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL092655.D</u>	CF 075 =	<u>PL092656.D</u>		
CF 050 =		<u>PL092657.D</u>	CF 025 =	<u>PL092658.D</u>	CF 005 =	<u>PL092659.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	2614880000	2506250000	2443430000	2489210000	2396540000	2490060000	3
4,4'-DDE	3334990000	3225880000	3154950000	3209960000	3182260000	3221610000	2
4,4'-DDT	2797500000	2713050000	2637310000	2653690000	2609750000	2682260000	3
Aldrin	3840860000	3705070000	3619180000	3619340000	3516600000	3660210000	3
alpha-BHC	4282500000	4089170000	3973860000	3934480000	3670850000	3990170000	6
alpha-Chlordane	3409310000	3303380000	3254510000	3329310000	3333630000	3326030000	2
beta-BHC	1625150000	1594840000	1591160000	1661990000	1758110000	1646250000	4
Decachlorobiphenyl	2606810000	2575500000	2605540000	2793460000	3064890000	2729240000	8
delta-BHC	4088000000	3924730000	3802680000	3782230000	3618150000	3843150000	5
Dieldrin	3483370000	3364390000	3290100000	3303460000	3260200000	3340300000	3
Endosulfan I	3111480000	3032780000	2993580000	3072340000	3055530000	3053140000	1
Endosulfan II	2881520000	2813750000	2779930000	2861240000	2829950000	2833280000	1
Endosulfan sulfate	2707530000	2646200000	2623690000	2712350000	2794620000	2696880000	2
Endrin	2969490000	2878380000	2828080000	2876210000	2912860000	2893010000	2
Endrin aldehyde	2273700000	2244250000	2230130000	2333460000	2454860000	2307280000	4
Endrin ketone	3089790000	3023040000	3013060000	3097240000	3120850000	3068800000	2
gamma-BHC (Lindane)	4083950000	3934430000	3833920000	3828430000	3616530000	3859450000	4
gamma-Chlordane	3470230000	3355710000	3294030000	3342020000	3346470000	3361690000	2
Heptachlor	3876200000	3766580000	3709120000	3779090000	3738650000	3773930000	2
Heptachlor epoxide	3405420000	3318630000	3272090000	3352830000	3358060000	3341410000	1
Methoxychlor	1400820000	1385450000	1393920000	1470360000	1489590000	1428030000	3
Tetrachloro-m-xylene	2724750000	2661560000	2643180000	2728430000	2847900000	2721160000	3